



## Stack-gas Analysis System

### ENDA 5000 series

**NO<sub>x</sub>, SO<sub>2</sub>, CO, CO<sub>2</sub>, O<sub>2</sub>**  
Continuous simultaneous 5-component analysis

#### **C**OMPACT

Uses half the space of previous models.

#### **E**ASY

Features an intuitive touch panel.

#### **L**ONG-TERM STABILITY

Uses NDIR for better long-term stability and reliability.

Steam boilers

Iron and steel processing

Refuse incinerators

Electric power generation plants

Sulfuric acid plants


Glass furnaces

		RANGE	
NO <sub>x</sub>	128.1 ppm	200	
SO <sub>2</sub>	120.2 ppm	200	
CO	153.2 ppm	200	
CO <sub>2</sub>	4.135 vol%	5	
O <sub>2</sub>	9.34 vol%	10	

MEAS. JAN/13/2004 15:21


MENU COR. CAL.

# The ENDA-5000 series of stack-gas analysis systems



## Continuous simultaneous and high-precision measurement of NOx, SO<sub>2</sub>, CO, CO<sub>2</sub>, and O<sub>2</sub>



Over 100,000 systems installed and 30 years of quality and experience. That is the base on which HORIBA's ENDA-5000 series of stack-gas analysis systems is built. These systems have a smaller footprint, and use cross-flow modulated non-dispersive infrared (NDIR) detection with a magnet-pneumatic detection method that is inherently drift-free. The ENDA-5000 series are superior continuous analysis systems that are perfect in the difficult field of exhaust gas measurement, where measurement errors cannot be tolerated. The series features an intuitive touch panel that makes every operation available with the touch of a single button. The ENDA-5000 series are also designed for faster, more efficient maintenance. They are ideal for a variety of uses, including monitoring steam boiler, refuse incinerator, and electric power generation plant emissions to assure pollution standards are being met.



## EASY

Features an intuitive touch panel.

## Easy to use



Measured concentration

Converted concentration



Correction history

Alarm history

The ENDA-5000 series use a large-format LCD touch panel that can display all five critical components (NOx, SO<sub>2</sub>, CO, CO<sub>2</sub>, O<sub>2</sub>) simultaneously. The touch panel also allows the operator to view the density variations of multiple components at once.


The operator can easily switch between the corrected and converted density settings screens or view alert information with the touch of a single button.

## COMPACT

Body yields wider maintenance area  
**Compact** (all maintenance can be done from the front)


ENDA-5000 series is a total analysis system which each equipment is housed in a 30cm depth cabinet (3-cylinder type). Therefore, ENDA-5000 series can be installed almost anywhere, with ample room on all sides for easy access and much easier maintenance. The compact blowback panel for sampling is available as well.

Downsizing of these equipments helps save space even when permanently installed, and free up valuable floor area for other equipment.



### Continuous simultaneous measurement of up to five components with one system

HORIBA'S innovative optical technology enables ENDA-5000 series measure up to five components which can be arranged any combination.



### Correction for interference

The interference correcting sensor uses a unique interference filter to compensate for the influence of interference by other gases.




### Better alerts and extra alerts

In addition to the alert functions available in the past, the ENDA-5000 series feature extra alert functions. A continuous checking process can prevent the unit from stopping due to a failure, reducing the risk of failed measurements and assuring consistent operation. US Patent No. 5,966,676



### Dramatically reduced correction time for SO<sub>2</sub>

Corrections of SO<sub>2</sub> measurements using wet base methods of the past took a great deal of time (about 15 minutes), but with the ENDA-5000 series' dry base method, correction takes only three minutes.



# The ultimate in dependability and reliability

## LONG-TERM STABILITY

Cross-flow modulated non-dispersive infrared (NDIR) detection is renowned for long-term stability.

### Long-term stability

#### 1 No need for optical adjustments

With cross-flow modulated non-dispersive infrared (NDIR) detection, the sample gas and reference gas are intrmitted into a single measurement cell alternately to obtain modulation signal. Therefore there is no need to adjust two different optical paths so that they are balanced.

#### 2 A stable zero point

Since the ENDA-5000 series output the difference between the measured gas and the reference gas each time measurement occurs (once a second), the zero point is extremely stable.

#### 3 Continuous cleaning keeps the cell clean


Since cleaning air is fed into the sample cell in between each batch of sample gas, the cell resists contamination and normally remains clean. This reduces span drift and makes the equipment safe and stable for long periods of time.

#### Other merits

- A CO<sub>2</sub> sensor constantly measures and makes corrections to compensate for CO<sub>2</sub> interference in NO<sub>x</sub> measurements.
- An interference compensation detector compensates for interference from H<sub>2</sub>O during NO<sub>x</sub> and SO<sub>2</sub> measurement.

**HORIBA's NDIR method, trusted through years of experience**


■ Cross-flow modulated non-dispersive infrared detection



#### ( Dual optical path comparison )

The sample gas and reference gas are each put in a different cell for measurement.


- This method is appropriate for laboratory use and other kinds of batch measurement.
- Differences in the cleanliness of the two optical paths result in signal variation.
- Keeping the cells clean requires periodic cleaning.
- During transport, and inspection, and whenever the unit is subject to vibrations, fine adjustment by a skilled service person is required.



■ Two optical path measurement

With magneto-pneumatic detection, there is no need for cylinder carrier gas.

The ENDA-5000 series use magneto-pneumatic detection to measure O<sub>2</sub>. Since the sample gas does not come into direct contact with the detector, there is no deterioration due to corrosion, which enables long-term stable operation. What's more, thanks to HORIBA's innovative technology, in which ambient air is used as a carrier gas, there is no need for a carrier gas supply, which translates into lower costs.



■ Measurement principle of magneto-pneumatic detection


■ Continuous correction is provided by a sensor that is designed to detect CO<sub>2</sub> interference during NO<sub>x</sub> measurement.

■ The systems feature an automatic recalibration function that calibrates the system every seven days.

■ A variety of types functions (up to 12 kinds of output)

- Instantaneous output (NO<sub>x</sub>, SO<sub>2</sub>, CO, CO<sub>2</sub>, O<sub>2</sub>)
- O<sub>2</sub> calculated output values (NO<sub>2</sub>, SO<sub>2</sub>, CO)
- Moving average values (for one to four hours)

■ Ambient air is used as the carrier gas, which allows for installation in smaller spaces and lower running costs.



■ Environmentally friendly thanks to lower electrical draw

These systems use 25% less electricity (200 VA) than older similar models.



# offers options for a variety of uses, all based on HORIBA

## Sampling sections

The ENDA-5000 series' sampling sections use cost-effective parts for maintenance, and offer a variety of sample gas conditioning systems, each suitable for a different kind of gas. HORIBA's know-how has created the best possible system for every type of sample gas measurement.

- Sample gas probe with easy-to-change filter element



Sample gas probe

- An innovative dehumidifying system minimizes loss of soluble components.
- A mist catcher in the sample flow path removes SO<sub>3</sub> and prevents damage and line blockage.
- Long-lasting, low-temperature (180°C) NO<sub>2</sub> → NO converter prevents corrosion.

### New pressure control \*

The new pressure control method is compatible with Daily start-up and shut-down and other intermittent operations.

\* Older models used a water filled pressure trap.

### Blowback panel reduced in size




In the past, a large blowback panel was necessary to control dust when measuring high-dust gas samples. HORIBA has used its innovative technology to reduce the size of the blowback panel by almost 25% (to 350 [W] x 550 [H] x 180 [D] mm). The panel is also lighter, and can be mounted on a wall. The new blowback panel can be used even in extremely small spaces.



## Models and components measured

NO <sub>x</sub>	SO <sub>2</sub>	CO <sub>2</sub>	CO	O <sub>2</sub>	Model
●					ENDA-5120
	●				ENDA-5130
		●			ENDA-5140
			●		ENDA-5150
				●	ENDA-5160
●				●	ENDA-5220
	●			●	ENDA-5230
		●		●	ENDA-5240
			●	●	ENDA-5250
●	●				ENDA-5300
●		●			ENDA-5310
●			●		ENDA-5320
	●	●			ENDA-5340
	●		●		ENDA-5350
		●	●		ENDA-5370
●	●			●	ENDA-5400
●		●		●	ENDA-5410
●			●	●	ENDA-5420
	●	●		●	ENDA-5440
	●		●	●	ENDA-5450
		●	●	●	ENDA-5470
●	●	●			ENDA-5500
●	●		●		ENDA-5510
●		●	●		ENDA-5520
	●	●	●		ENDA-5530
●	●	●		●	ENDA-5600
●	●		●	●	ENDA-5610
●		●	●	●	ENDA-5620
	●	●	●	●	ENDA-5630
●	●	●	●		ENDA-5700
●	●	●	●	●	ENDA-5800

# A's extensive know-how.



\*Systems can be customized for a variety of other applications as well.

## Specifications

Model		ENDA-5000				
Component		NO <sub>x</sub>	SO <sub>2</sub>	CO *1	CO <sub>2</sub>	O <sub>2</sub> *2
Measurement methods		NDIR	NDIR	NDIR	NDIR	Magneto-pneumatic detection
Range	Standard	200~5000 ppm	200~5000 ppm	200~5000 ppm	5~25 vol%	10~25 vol%
	Option	100 ppm~	50 ppm~	100 ppm~	—	—
Range Ratio		Within a factor of 10	Within a factor of 10	Within a factor of 10	Within a factor of 5	Within a factor of 2.5
Repeatability		Within 0.5% of full scale (with optional range, or during O <sub>2</sub> measurement, ±1.0% of full scale)				
Linearity (indicator error)		± 1.0% of full scale				
Zero drift		± 1.0% of full scale/week (assuming surrounding temperature is maintained within 5°C) (with optional range, or O <sub>2</sub> measurement, ± 2.0% of full scale/week)				
Span drift		± 2.0% of full scale/week (assuming surrounding temperature is maintained within 5°C)				
Response time		Within 60 seconds (T <sub>d</sub> + T <sub>90</sub> from equipment intake area) (sample flow 0.6 L/min.) (within 240 seconds for SO <sub>2</sub> only)				
Interference		± 2.0% of full scale (within standard range, with standard gas formation)				
Display		Touch panel LCD (backlight) (four usable lines)				
Environment Condition	Temperature	-5 to 40°C (away from direct sunlight and radiation heat) *3				
	Humidity	90% or less (no condensation)				
	Vibration	100 Hz, 0.3 m/s <sup>2</sup> or less				
	Dust	Standard environment or better				
Measuring Gas Condition	Temperature	250°C or lower				
	Dust	0.1 g/Nm <sup>3</sup> or less				
Standard gas composition *4		NO: 500 ppm or less; NO <sub>2</sub> : 15 ppm or less; SO <sub>2</sub> : 1000 ppm or less; SO <sub>3</sub> : 50 ppm or less; CO: 200 ppm or less; CO <sub>2</sub> : 15 vol% or less; H <sub>2</sub> O: 20 vol% or less				
Sampling method		Dry sampling using an electric cooler				
Sample gas flow		2.5 L/min~3.0 L/min				
Sample inlet tube		PTFE tubing (ø8/ø6 mm)				
Sample gas pressure		± 4.9 kPa (three points selected) (with no sample gas back pressure)		(1) -1.96 to 4.9 kPa (2) ± 3.43 kPa (3) -4.9 to 1.96 kPa		
	Pressure control	Pressure control uses a regulator and pump; Reduced pressure sampling; Control pressure: -4.9 kPa				
	Output	DC 4 to 20 mA (absolute output) (DC 0 to 16 mA/DC 0 to 1 V/DC 1 to 5V optional) Max. 12 output systems				
External output		Analysis alerts, analysis warnings, range display, corrections, conservation, purging (option) Contact capacity: DC 30 V 1 A, AC 250 V 1 A resistance load				
Correction method		Dry correction, automatic correction (correction cycle: 7 days standard, can be adjusted to between 1 and 99 days), manual correction				
Calibration gas	Zero gas	With measurement method authorization: N <sub>2</sub> . When there is no measurement method authorization: N <sub>2</sub> or ambient air				
	O <sub>2</sub> carrier gas	Ambient air				
	Span gas	Gas cylinder for each component measured (when there is no measurement method authorization: O <sub>2</sub> or ambient air can be used)				
Probe		Flange: JIS 10K, 40 AFF; Sample probe tube length: 1000 mm; Material: SUS-316 stainless steel;				
Primary filter		Filter element: SUS-304 stainless steel and 2µm-pleated quartz wool; Electric heater: 100 VA, with water droplet proof case				
Power supply		AC 100 V ± 15 V (85 V ~ 115 V)				
Power frequency		50/60 Hz (switchable)				
Power consumption		About 800 VA (heated piping 30m: +1100 VA; heater in tray: +300 VA)				
Exterior dimensions / Mass		600 (W) x 1770 (H) x 300 (D) mm (high pressure gas cylinders, 3.4 L cylinders, maximum of 3 cylinders); About 180 kg (not including cylinders)				
		600 (W) x 1770 (H) x 500 (D) mm (high pressure gas cylinders, 3.4 L cylinders, maximum of 6 cylinders); About 200 kg (not including cylinders)				
Materials in contact with sample gas		SUS-316 stainless steel, SUS-304 stainless steel, PTFE, polypropylene, polyethylene, fluororubber, PVC, PVDF, and glass				
Enclosure		Independent outdoor installation Plate thickness: Main unit, doors, top plate, steel plate: 2.3 mm; Channel base: 3.2 mm; Doors: front opening; Interface: right front				
Color/Finish		Semi-gloss Munsell 5Y7/1 on all inner and outer surfaces				

\*1: The analyzer against N<sub>2</sub>O interference for CO analyzer applies the standard range of 200 ppm or more (no optional range).


\*2: No carrier gas cylinder is necessary.

\*3: For the temperature range of -15°C to 40°C (cold district) and for the specification of -5°C to 50°C, we will separately discuss the design.

\*4: When the coexisting gas contains NH<sub>3</sub>, NH<sub>3</sub> scrubber is prepared for an optional part.

When CH<sub>4</sub> coexists in the sample gas for SO<sub>2</sub> analyzer, the SO<sub>2</sub> analyzer of CH<sub>4</sub> interference compensation type is prepared. When N<sub>2</sub>O coexists in the sample gas for CO analyzer, the CO analyzer of N<sub>2</sub>O interference compensation type is prepared.

## Dimensions (unit: mm)



The EMC Directive : EN61326 Compliant  
European Standard compliant : EN15267, EN14181  
Pattern Approved, Metrology Law (China) : 2006-C118



The HORIBA Group adopts IMS (Integrated Management System) which integrates Quality Management System ISO9001, Environmental Management System ISO14001, and Occupational Health and Safety Management System OHSAS18001. We have now integrated Business Continuity Management System ISO22301 in order to provide our products and services in a stable manner, even in emergencies.



Please read the operation manual before using this product to assure safe and proper handling of the product.

- The contents of this catalog are subject to change without prior notice, and without any subsequent liability to this company.
- The color of the actual products may differ from the color pictured in this catalog due to printing limitations.
- It is strictly forbidden to copy the content of this catalog in part or in full.
- All brand names, product names and service names in this catalog are trademarks or registered trademarks of their respective companies.

http://www.horiba.com e-mail: info@horiba.co.jp

● **HORIBA, Ltd.**  
Head Office  
2 Miyano Higashi, Kisshoin,  
Minami-ku, Kyoto, Japan  
Phone: 81 (75) 313-8121  
Fax: 81 (75) 321-5725

● **HORIBA (China) Trading Co., Ltd.**  
Shanghai  
Unit D, 1F, Building A, Synnex  
International Park, 1068 West  
Tianshan Road, Shanghai,  
200335, China  
Phone: 86 (21) 6289-6060  
Fax: 86 (21) 6289-5553

● **HORIBA Korea Ltd.**  
East Office  
10, Dogok-Ro, 6-Gil,  
Gangnam-Gu, Seoul-Si,  
06259, Korea  
Phone: 82 (2) 753-7911  
Fax: 82 (2) 756-4972

● **HORIBA (Thailand) Ltd.**  
East Office  
850 / 7 Soi Lat Krabang  
30 / 5, Lat Krabang Road,  
Lat Krabang, Bangkok  
10520, Thailand  
Phone: 66 (0) 2734 4434  
Fax: 66 (0) 2734 4438

● **HORIBA Instruments Incorporated**  
Irvine Office  
9755 Research Drive, Irvine,  
CA 92618, U.S.A.  
Phone: 1 (949) 250-4811  
Fax: 1 (949) 250-0924

● **HORIBA Instruments**  
Alvin, TX  
5318 W.FM517 Rd, Alvin,  
TX 77511, U.S.A.  
Phone: 1 (281) 482-4334  
Fax: 1 (281) 614-0303

● **HORIBA Instruments**  
(Singapore) Pte Ltd.  
Head Office  
3 Changi Business Park  
Vista #01-01 Akzonobel  
House, Singapore 486051  
Phone: 65 (6) 745-8300  
Fax: 65 (6) 745-8155

● **HORIBA Vietnam Co., Ltd.**  
Unit 6, 10 Floor, CMC Tower,  
Duy Tan Street, Dich Vong  
Hau Ward, Cau Giay District,  
Hanoi, Vietnam  
Phone: 84 (4) 3795-8552  
Fax: 84 (4) 3795-8553

● **PT HORIBA Indonesia**  
Jl. Jalur Sutera Blok 20A,  
No.16-17, Kel. Kunciran, Kec.  
Pinang Tangerang-15144,  
Indonesia  
Phone: 62 (21) 3044-8525  
Fax: 62 (21) 3044-8521

● **HORIBA India Private Limited**  
Delhi  
246, Okhla Industrial Estate,  
Phase 3 New Delhi-110020,  
India  
Phone: 91 (11) 4646-5000  
Fax: 91 (11) 4646-5020

● **HORIBA India Private Limited**  
Pune  
502, 5th Floor, Purushottam  
Plaza, Baner Road, Baner,  
Pune-411045 India  
Phone: 91 (20) 4076-6000  
Fax: 91 (20) 4076-6010

● **HORIBA Instruments**  
Bangalore  
Kamadhenu, No.17 / 1 - 32,  
Bannerghatta Road, Audugodi  
Bangalore-560030 India  
Phone: 91 (80) 22210071

● **HORIBA Instruments**  
Brasil, Ltda.  
Head Office  
Rua: Presbitero Plinio Alves  
de Souza, 645, Loteamento  
Polo Multivas Barreiro  
Medeiros-Jundiai Sao Paulo  
CEP 13.212-181 Brazil  
Phone: 55 (11) 2923-5400  
Fax: 55 (11) 2923-5490

● **HORIBA UK Limited**  
Northampton  
Kyoto Close Moulton Park,  
Northampton NN3 6FL, UK  
Phone: 44 (1604) 542-500  
Fax: 44 (1604) 542-699

● **HORIBA (Austria) GmbH**  
Kaplanstrasse 5 A-3430  
Tulln, Austria  
Phone: 43 (2272) 65225  
Fax: 43 (2272) 65230

● **HORIBA Czech**  
Prumyslova 1306 / 7,  
CZ-10200, Praha 10,  
Czech Republic  
Phone: 420 (2) 460-392-65

● **HORIBA Europe GmbH**  
Head Office  
Hans-Mess-Str.6 D-61440  
Oberursel Germany  
Phone: 49 (6172) 1396-0  
Fax: 49 (6172) 1373-85

● **Leichlingen Office**  
Julius-kronenberg Str.9  
D-42799 Leichlingen  
Germany  
Phone: 49 (2175) 8978-0  
Fax: 49 (2175) 8978-50

● **HORIBA France Sarl**  
12, Av des Tropiques Hightec  
Sud, F-91955 Les Ulis, France  
Phone: 33 (1) 69-29-96-23  
Fax: 33 (1) 69-29-95-77

# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

## **ENDA-5000 Series Stack Gas Analysis System**

Manufactured by:

### **Horiba GmbH**

Kaplanstrasse 5  
A-3430 Tulln  
Austria

Has been assessed by Sira Certification Service  
And for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Emission  
Monitoring Systems, Version 3.5 dated June 2016  
EN15267-3:2007,  
& QAL 1 as defined in EN 14181: 2004**

Certification Ranges :

CO	0 to 50 mg/m <sup>3</sup>	0 to 500 mg/m <sup>3</sup>
NO <sub>x</sub>	0 to 100 mg/m <sup>3</sup>	0 to 1000 mg/m <sup>3</sup>
SO <sub>2</sub>	0 to 75 mg/m <sup>3</sup>	0 to 750 mg/m <sup>3</sup>
CO <sub>2</sub>	0 to 20 Vol.-%	
O <sub>2</sub>	0 to 25 Vol.-%	

Project No.: 16A28926/7017426  
Certificate No: Sira MC 120212/02  
Initial Certification: 26 November 2012  
This Certificate issued: 28 February 2018  
Renewal Date: 25 November 2022

Joe Prince MSc, MInst MC  
Certification Manager

MCERTS is operated on behalf of the Environment Agency by

## **Sira Certification Service**

Unit 6, Hawarden Industrial Park  
Hawarden, Deeside, CH5 3US  
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.  
For conditions of use, please consider all the information within.*

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at [www.mcerts.net](http://www.mcerts.net)

On the basis of the assessment and the ranges required for compliance with EU Directives this instrument is considered suitable for use on waste incineration and large coal-fired combustion plant applications. This CEM has been proven suitable for its measuring task (parameter and composition of the flue gas) by use of the QAL 1 procedure specified in EN14181, for LCPD and WID applications for the ranges specified. The lowest certified range for each determinand shall not be more than 1.5X the daily average emission limit value (ELV) for WID applications, and not more than 2.5X the ELV for LCPD and other types of application.

The field test was conducted on a municipal waste incineration plant downstream of the exhaust gas cleaning system for 8 months, between 13<sup>th</sup> May 2011 and 26<sup>th</sup> January 2012.


## Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

- TUV Report No. 936/21218525/A\_en dated 20/03/2012
- QAL 1 Test Report for HORIBA ENDA 5000 Multi Gas Analyser – Corrected Tables for Lack-of-fit test in the field.

## Product Certified

The Horiba ENDA-5000 series Stack Gas Analysis System consists of the following parts:



<b>1. Sample Probe</b>	<b>2. Heated Filter</b>	<b>3. Heated Sample Line</b>	<b>4. Gas Conditioning</b>	<b>5. Analyser</b>
Model: M&C PSP 4000-H	Model: N/A - Integrated with sampling probe	Model: RACO Sample Line Length: 10m	Model: ENDA-5000	Model: ENDA-5000 with CMA-5800 Analysing module

This certificate applies to all instruments fitted with software version P10008770011 (serial number 0900500 onwards).

Certificate No : Sira MC120212/02  
This Certificate issued : 28 February 2018

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*



## Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: 5°C to 40°C  
Instrument IP rating: IP54

Results are expressed as % of the certification range, unless otherwise stated. O<sub>2</sub> results are reported as %vol.

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Response time						
NO <sub>x</sub>					60s	<200s
SO <sub>2</sub>					117s	<200s
CO					64s	<200s
CO <sub>2</sub>					64s	<200s
O <sub>2</sub>					57s	<200s
Repeatability standard deviation at zero point						
NO <sub>x</sub>	0.1					<2.0%
SO <sub>2</sub>	0.1					<2.0%
CO		0.6				<2.0%
CO <sub>2</sub>	0.0					<2.0%
O <sub>2</sub>	0.01					<0.2%
Repeatability standard deviation at reference point						
NO <sub>x</sub>	0.2					<2.0%
SO <sub>2</sub>	0.2					<2.0%
CO			1.3			<2.0%
CO <sub>2</sub>	1.0					<2.0%
O <sub>2</sub>	0.02					<0.2%

Certificate No : Sira MC120212/02  
This Certificate issued : 28 February 2018

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Lack-of-fit						
NO <sub>x</sub>	0.3					<2.0%
SO <sub>2</sub>		0.8				<2.0%
CO		0.6				<2.0%
CO <sub>2</sub>			1.0			<2.0%
O <sub>2</sub>	0.07					<0.2%
Influence of ambient temperature zero point						
NO <sub>x</sub>		-0.5				<5.0%
SO <sub>2</sub>				-2.3		<5.0%
CO		0.5				<5.0%
CO <sub>2</sub>	-0.2					<5.0%
O <sub>2</sub>	0.43					<0.5%
Influence of ambient temperature reference point						
NO <sub>x</sub>			1.9			<5.0%
SO <sub>2</sub>				-2.0		<5.0%
CO				2.4		<5.0%
CO <sub>2</sub>			1.0			<5.0%
O <sub>2</sub>	0.11					<0.5%

Certificate No : Sira MC120212/02  
 This Certificate issued : 28 February 2018

*This certificate may only be reproduced in its entirety and without change  
 To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Influence of sample gas flow for extractive CEMS						
NO <sub>x</sub>	0.4					<2.0%
SO <sub>2</sub>	0.1					<2.0%
CO	-0.1					<2.0%
CO <sub>2</sub>	0.1					<2.0%
O <sub>2</sub>	0.07					<0.2%
Influence of voltage variations 190 to 250V						
NO <sub>x</sub>		-0.9				<2.0%
SO <sub>2</sub>			-1.0			<2.0%
CO		0.5				<2.0%
CO <sub>2</sub>		-0.6				<2.0%
O <sub>2</sub>	0.04					<0.2%
Influence of vibration (10 to 60Hz (±0.3mm), 60 to 150Hz at 19.6m/s <sup>2</sup> )					Not applicable for extractive CEMS.	To be reported
Cross-sensitivity at zero with interferents: O <sub>2</sub> , H <sub>2</sub> O, CO, CO <sub>2</sub> , CH <sub>4</sub> , N <sub>2</sub> O, NO, NO <sub>2</sub> , NH <sub>3</sub> , SO <sub>2</sub> , HCl						
NO <sub>x</sub>			1.96			<4.0%
SO <sub>2</sub>				3.80		<4.0%
CO				2.46		<4.0%
CO <sub>2</sub>	<0.5					<4.0%
O <sub>2</sub>	<0.1					<0.4%
Cross-sensitivity at reference with interferents: O <sub>2</sub> , H <sub>2</sub> O, CO, CO <sub>2</sub> , CH <sub>4</sub> , N <sub>2</sub> O, NO, NO <sub>2</sub> , NH <sub>3</sub> , SO <sub>2</sub> , HCl						
NO <sub>x</sub>				3.00		<4.0%
SO <sub>2</sub>				3.73		<4.0%
CO				2.60		<4.0%
CO <sub>2</sub>		-0.95				<4.0%
O <sub>2</sub>	-0.19					<0.4%
Converter Efficiency					97.6%	>95%

Certificate No : Sira MC120212/02  
 This Certificate issued : 28 February 2018

*This certificate may only be reproduced in its entirety and without change  
 To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Measurement uncertainty						
NO <sub>x</sub>					5.1%	Guidance - at least 25% below max permissible uncertainty
SO <sub>2</sub>					8.8%	
CO					5.6%	
CO <sub>2</sub>					4.1%	
O <sub>2</sub>					2.3%	
Calibration function (field)						
NO <sub>x</sub>					0.9967	>0.90
SO <sub>2</sub>					0.9960	>0.90
CO					0.9148	>0.90
CO <sub>2</sub>					0.9788	>0.90
O <sub>2</sub>					0.9994	>0.90
Response time (field)						
NO <sub>x</sub>					67s	<200s
SO <sub>2</sub>					98s	<200s
CO					54s	<200s
CO <sub>2</sub>					58s	<200s
O <sub>2</sub>					57s	<200s
Lack of fit (field)						
NO <sub>x</sub>	0.3					<2.0%
SO <sub>2</sub>	-0.47					<2.0%
CO	0.4					<2.0%
CO <sub>2</sub>			-1.8			<2.0%
O <sub>2</sub>	0.1					<0.2%
Maintenance interval					3 Months	Note 1 >8 days

Certificate No : Sira MC120212/02  
This Certificate issued : 28 February 2018

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*



Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Zero and Span drift requirement	<p>The Horiba ENDA-5000 system is an extractive system that uses various particulate filters to protect the optical system from contamination. The recommendations of the manufacturer with regard to filter changes should be followed. No compensation for zero and span drift is employed. Drift can be determined from normal automatic or manual zero and span checks.</p>					<p>Clause 6.13 &amp; 10.13</p> <p>Manufacturer shall provide a description of the technique to determine and compensate for zero and span drift.</p>
Change in zero point over maintenance interval						
NO <sub>x</sub>		-0.94				<3.0%
SO <sub>2</sub>			-1.44			<3.0%
CO		0.50				<3.0%
CO <sub>2</sub>		-0.63				<3.0%
O <sub>2</sub>	0.19					<0.2%
Change in reference point over maintenance interval						
NO <sub>x</sub>				-2.81		<3.0%
SO <sub>2</sub>			1.81			<3.0%
CO				2.25		<3.0%
CO <sub>2</sub>				2.69		<3.0%
O <sub>2</sub>	0.16					<0.2%
Availability					99.1%	>95% (>98% for O <sub>2</sub> )
Reproducibility						
NO <sub>x</sub>		0.7				<3.3%
SO <sub>2</sub>			1.1			<3.3%
CO			1.7			<3.3%
CO <sub>2</sub>		0.9				<3.3%
O <sub>2</sub>	0.19					<0.2%

Note 1: The ENDA 5000 Series Gas Analyser has a maintenance interval of 3 months. The work detailed below has to be carried out at regular intervals, depending on local conditions:

QAL3 activities in accordance with EN 14181 require monthly zero and span point checks for the component CO<sub>2</sub>. It has been noted that container volumes for phosphoric acid collectors need to be adjusted for a three months maintenance

Certificate No : Sira MC120212/02  
This Certificate issued : 28 February 2018

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

interval. It should be at least 120L. If smaller containers are used, these should be checked more often and refilled as the case may be.

### Description

The Horiba ENDA-5000 system is an extractive emission analysis system. This comprises a heated sample probe and filter, heated line, sample conditioning system and an analyser module for the measurement of up to five components. Carbon Monoxide, Carbon Dioxide, Oxides of Nitrogen, Sulphur Dioxide are measured by non-dispersive infra-red absorption (NDIR) and Oxygen by magneto-pneumatic method which does not require a carrier gas from a gas cylinder.

The sample conditioning system and analyser module are contained in a single unit with controller for the heated sample line.

### General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC 120212/00
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC120212/02  
This Certificate issued : 28 February 2018

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*